

Notice of Allowability

Application No.

10/070,924

Examiner

George R. Koch III

Applicant(s)

MATSUMOTO, TAKUYA

Art Unit

1734

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 12/23/2003.
2. ☒ The allowed claim(s) is/are 14-27.
3. ☒ The drawings filed on 23 December 2003 are accepted by the Examiner.
4. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☒ All b) ☐ Some* c) ☐ None of the:
 1. ☒ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

5. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 6. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
7. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

- | | |
|---|--|
| 1. <input type="checkbox"/> Notice of References Cited (PTO-892) | 5. <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 2. <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 6. <input type="checkbox"/> Interview Summary (PTO-413),
Paper No./Mail Date _____. |
| 3. <input type="checkbox"/> Information Disclosure Statements (PTO-1449 or PTO/SB/08),
Paper No./Mail Date _____ | 7. <input type="checkbox"/> Examiner's Amendment/Comment |
| 4. <input type="checkbox"/> Examiner's Comment Regarding Requirement for Deposit
of Biological Material | 8. <input checked="" type="checkbox"/> Examiner's Statement of Reasons for Allowance |
| | 9. <input type="checkbox"/> Other _____. |

DETAILED ACTION

Allowable Subject Matter

1. Claims 14-27 are allowed.
2. The following is an examiner's statement of reasons for allowance: This invention is deemed directed to a non-obvious improvement of the prior art such as the admitted prior art (Figure 1) and the Burtin references (US Patents 5,703,770 or 5,939,993). The improvement consists of the addition of a current sensor for detecting a current value of an intrinsic consumed current at the high voltage booster circuit. While current sensors are generically provided, the prior art does not suggest placing a sensor at a location for detecting a current value of an intrinsic consumed current at the high voltage booster circuit.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Information Disclosure Statement

3. The information disclosure statement filed 12/23/2003 fails to comply with 37 CFR 1.98(a)(1), which requires a list of all patents, publications, or other information submitted for consideration by the Office. It has been placed in the application file, but the information referred to therein has not been considered.

Drawings

4. The drawings correcting Figure 1 were received on 12/23/2003. These drawings are approved.
5. Applicant's remarks filed 12/23/2004 which indicated Figure 2 should not have a prior art label are considered persuasive. This requirement has been lifted.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to George R. Koch III whose telephone number is (571) 272-1230 (TDD only). If the applicant cannot make a direct TDD-to-TDD call, the applicant can communicate by calling the Federal Relay Service at 1-800-877-8339 and giving the operator the above TDD number. The examiner can normally be reached on M-Th 10-7.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Richard Crispino can be reached on (571) 272-1226. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 1734

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



George R. Koch III
March 7, 2004



RICHARD CRISPINO
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 1700